

Download Hyfrecator 2000 Service Manual

Hyfrecator® 2000 Electrosurgical Unit (ESU) up to original factory specifications. CONMED Corporation reserves the right to alter the specifications of the Hyfrecator® 2000. It is important to have access to the Operator's Manual supplied with each unit in order to restore original specifications. 2000 Section 9 - Maintenance The HYFRECACTOR® 2000 should be tested by qualified service personnel on a periodic basis. CONMED suggests examination of the unit at least every 12 months. NOTE: All internal service should be performed by qualified service personnel only. Page 24: Optional Accessories Conmed - Hyfrecator 2000; Documents; Service Manual; Conmed Hyfrecator 2000 Service Manual. Conmed - Hyfrecator 2000 by Conmed. Download PDF. Product Details. Forums; Documents; Parts; Videos; News; Request a quote Request Parts Add to My Bench. Conmed Hyfrecator 2000 Service Manual in English, 7-900-SM-ENG. Guides the qualified biomedical electronics technician through all aspects of preventative maintenance and repairs necessary to keep the Hyfrectaor 2000 up to original factory specifications. NOTE: We are an authorized distributor of Conmed Electrosurgical Equipment and Accessories. We carry the full line of Conmed Hyfrecator 2000, parts and accessories.